



CONDUCTIVITY STANDARD INFORMATION

Equipment

- 1413 μS Conductivity Standard
- For calibration of conductivity meters. Contains 0.01M KCl, preservative
- Store away from direct sunlight at 4-25°C
- Replace cap after use
- Do not return reagent to bottle

Temperature/ μS Chart	
$^{\circ}\text{C}$	μS
0	779
5	896
10	1020
15	1147
20	1278
25	1413
30	1512